

16220  
16792



## Declaration of conformity

According to the MDD/93/42/EEC, Medical Devices Directive.

# PALETTO PLUS and PALETTO WALL PLUS

Article no: 100090 and 100193

**Type of equipment:** Communication training device

**Manufacturer:** KOMIKAPPAB, Stormhallsvägen 4, SE-432 32 Varberg, SWEDEN

**Brand name and trademark:** KOMIKAPP, PALETTO

Conformity with the Directive stated above relates to the following reference documents:

EN 71:1  
Test report: Mechanical and physical properties  
Statens Provningsanstalt, Borås

EN 71:3  
Test report: Migration of certain elements  
Statens Provningsanstalt, Borås

EN 60601-1-2:2001  
Test reports: Electromagnetic Compatibility  
Nemko, Norway

EN 50081-2-2:1993,  
Test reports: Generic Emission Standard, Industry  
Telecom Lund

EN 55022 1.1  
Test reports: Emission enclosure port  
Telecom Lund

Transformer:  
IEC 60601-1 Low Voltage Directive 73/23/EEC

EN 61000-6-2 ; EN 61000-6-3 EMC Directives 89/336/EEC, 92/31/EEC

As manufacturer established within EC, we declare under our sole responsibility that the equipment follows the provisions of the directives stated above.

2014-01-01

.....  
Pernilla Karlsson  
General Manager